

November 27, 2019

Shane Grant
Eastern Water Solutions
5 Benson Road
Oxford, CT 06478

Project Location: 4 Barbourtown Rd. Canton, CT
Client Job Number:
Project Number: WELL #1, Y2hp CT 0230392
Laboratory Work Order Number: 19K0630

Enclosed are results of analyses for samples received by the laboratory on November 11, 2019. If you have any questions concerning this report, please feel free to contact me.

Sincerely,

A handwritten signature in black ink that reads "Michelle Koch". The signature is written in a cursive style with a large initial "M" and a long, sweeping underline.

Michelle M. Koch
Project Manager

Table of Contents

Sample Summary	3
Case Narrative	4
Sample Results	5
19K0630-01	5
19K0630-02	6
Sample Preparation Information	7
QC Data	8
Semivolatile Organic Compounds by - LC/MS-MS	8
B246729	8
Flag/Qualifier Summary	9
Certifications	10
Chain of Custody/Sample Receipt	11

39 Spruce Street * East Longmeadow, MA 01028 * FAX 413/525-6405 * TEL. 413/525-2332

Eastern Water Solutions
5 Benson Road
Oxford, CT 06478
ATTN: Shane Grant

REPORT DATE: 11/27/2019

PURCHASE ORDER NUMBER:

PROJECT NUMBER: WELL #1, Y2hp CT 0230392

ANALYTICAL SUMMARY

WORK ORDER NUMBER: 19K0630

The results of analyses performed on the following samples submitted to the CON-TEST Analytical Laboratory are found in this report.

PROJECT LOCATION: 4 Barbourtown Rd. Canton, CT

FIELD SAMPLE #	LAB ID:	MATRIX	SAMPLE DESCRIPTION	TEST	SUB LAB
Well #1 Field Blank	19K0630-01	Drinking Water		EPA 537.1	
Well #1	19K0630-02	Drinking Water		EPA 537.1	

CASE NARRATIVE SUMMARY

All reported results are within defined laboratory quality control objectives unless listed below or otherwise qualified in this report.

The results of analyses reported only relate to samples submitted to the Con-Test Analytical Laboratory for testing.

I certify that the analyses listed above, unless specifically listed as subcontracted, if any, were performed under my direction according to the approved methodologies listed in this document, and that based upon my inquiry of those individuals immediately responsible for obtaining the information, the material contained in this report is, to the best of my knowledge and belief, accurate and complete.

A handwritten signature in black ink, appearing to read "Tod Kopycinski". The signature is written in a cursive style with a large, sweeping initial "T".

Tod E. Kopycinski
Laboratory Director

39 Spruce Street * East Longmeadow, MA 01028 * FAX 413/525-6405 * TEL. 413/525-2332

Project Location: 4 Barbourtown Rd. Canton, CT

Sample Description:

Work Order: 19K0630

Date Received: 11/11/2019

Field Sample #: Well #1 Field Blank

Sampled: 11/11/2019 12:00

Sample ID: 19K0630-01

Sample Matrix: Drinking Water

Semivolatile Organic Compounds by - LC/MS-MS

Analyte	Results	MCL/SMCL		Units	Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
		RL	MA ORSG							
Perfluorobutanesulfonic acid (PFBS)	ND	2.0		ng/L	1		EPA 537.1	11/22/19	11/27/19 4:02	BLM
Perfluorohexanoic acid (PFHxA)	ND	2.0		ng/L	1		EPA 537.1	11/22/19	11/27/19 4:02	BLM
Perfluorohexanesulfonic acid (PFHxS)	ND	2.0		ng/L	1		EPA 537.1	11/22/19	11/27/19 4:02	BLM
Perfluoroheptanoic acid (PFHpA)	ND	2.0		ng/L	1		EPA 537.1	11/22/19	11/27/19 4:02	BLM
Perfluorooctanoic acid (PFOA)	ND	2.0		ng/L	1		EPA 537.1	11/22/19	11/27/19 4:02	BLM
Perfluorooctanesulfonic acid (PFOS)	ND	2.0		ng/L	1		EPA 537.1	11/22/19	11/27/19 4:02	BLM
Perfluorononanoic acid (PFNA)	ND	2.0		ng/L	1		EPA 537.1	11/22/19	11/27/19 4:02	BLM
Perfluorodecanoic acid (PFDA)	ND	2.0		ng/L	1		EPA 537.1	11/22/19	11/27/19 4:02	BLM
N-EtFOSAA	ND	2.0		ng/L	1		EPA 537.1	11/22/19	11/27/19 4:02	BLM
Perfluoroundecanoic acid (PFUnA)	ND	2.0		ng/L	1		EPA 537.1	11/22/19	11/27/19 4:02	BLM
N-MeFOSAA	ND	2.0		ng/L	1		EPA 537.1	11/22/19	11/27/19 4:02	BLM
Perfluorododecanoic acid (PFDoA)	ND	2.0		ng/L	1		EPA 537.1	11/22/19	11/27/19 4:02	BLM
Perfluorotridecanoic acid (PFTrDA)	ND	2.0		ng/L	1		EPA 537.1	11/22/19	11/27/19 4:02	BLM
Perfluorotetradecanoic acid (PFTA)	ND	2.0		ng/L	1		EPA 537.1	11/22/19	11/27/19 4:02	BLM
Hexafluoropropylene oxide dimer acid (HFPO-DA)	ND	2.0		ng/L	1		EPA 537.1	11/22/19	11/27/19 4:02	BLM
11Cl-PF3OUdS (F53B Major)	ND	2.0		ng/L	1		EPA 537.1	11/22/19	11/27/19 4:02	BLM
9Cl-PF3ONS (F53B Minor)	ND	2.0		ng/L	1		EPA 537.1	11/22/19	11/27/19 4:02	BLM
4,8-dioxa-3H-perfluorononanoic acid (ADONA)	ND	2.0		ng/L	1		EPA 537.1	11/22/19	11/27/19 4:02	BLM
Surrogates		% Recovery		Recovery Limits		Flag/Qual				
13C-PFHxA		93.4		70-130					11/27/19 4:02	
M3HFPO-DA		89.3		70-130					11/27/19 4:02	
13C-PFDA		117		70-130					11/27/19 4:02	
d5-NEtFOSAA		119		70-130					11/27/19 4:02	

39 Spruce Street * East Longmeadow, MA 01028 * FAX 413/525-6405 * TEL. 413/525-2332

Project Location: 4 Barbourtown Rd. Canton, CT

Sample Description:

Work Order: 19K0630

Date Received: 11/11/2019

Field Sample #: Well #1

Sample ID: 19K0630-02

Start Date/Time: 11/11/2019 12:06:00PM

Sample Matrix: Drinking Water

Stop Date/Time: 11/11/2019 12:10:00PM

Semivolatile Organic Compounds by - LC/MS-MS

Analyte	Results	MCL/SMCL			Dilution	Flag/Qual	Method	Date Prepared	Date/Time Analyzed	Analyst
		RL	MA ORSG	Units						
Perfluorobutanesulfonic acid (PFBS)	ND	2.0		ng/L	1	EPA 537.1	11/22/19	11/27/19 4:24	BLM	
Perfluorohexanoic acid (PFHxA)	ND	2.0		ng/L	1	EPA 537.1	11/22/19	11/27/19 4:24	BLM	
Perfluorohexanesulfonic acid (PFHxS)	ND	2.0		ng/L	1	EPA 537.1	11/22/19	11/27/19 4:24	BLM	
Perfluoroheptanoic acid (PFHpA)	ND	2.0		ng/L	1	EPA 537.1	11/22/19	11/27/19 4:24	BLM	
Perfluorooctanoic acid (PFOA)	ND	2.0		ng/L	1	EPA 537.1	11/22/19	11/27/19 4:24	BLM	
Perfluorooctanesulfonic acid (PFOS)	ND	2.0		ng/L	1	EPA 537.1	11/22/19	11/27/19 4:24	BLM	
Perfluorononanoic acid (PFNA)	ND	2.0		ng/L	1	EPA 537.1	11/22/19	11/27/19 4:24	BLM	
Perfluorodecanoic acid (PFDA)	ND	2.0		ng/L	1	EPA 537.1	11/22/19	11/27/19 4:24	BLM	
N-EtFOSAA	ND	2.0		ng/L	1	EPA 537.1	11/22/19	11/27/19 4:24	BLM	
Perfluoroundecanoic acid (PFUnA)	ND	2.0		ng/L	1	EPA 537.1	11/22/19	11/27/19 4:24	BLM	
N-MeFOSAA	ND	2.0		ng/L	1	EPA 537.1	11/22/19	11/27/19 4:24	BLM	
Perfluorododecanoic acid (PFDoA)	ND	2.0		ng/L	1	EPA 537.1	11/22/19	11/27/19 4:24	BLM	
Perfluorotridecanoic acid (PFTrDA)	ND	2.0		ng/L	1	EPA 537.1	11/22/19	11/27/19 4:24	BLM	
Perfluorotetradecanoic acid (PFTA)	ND	2.0		ng/L	1	EPA 537.1	11/22/19	11/27/19 4:24	BLM	
Hexafluoropropylene oxide dimer acid (HFPO-DA)	ND	2.0		ng/L	1	EPA 537.1	11/22/19	11/27/19 4:24	BLM	
11Cl-PF3OUdS (F53B Major)	ND	2.0		ng/L	1	EPA 537.1	11/22/19	11/27/19 4:24	BLM	
9Cl-PF3ONS (F53B Minor)	ND	2.0		ng/L	1	EPA 537.1	11/22/19	11/27/19 4:24	BLM	
4,8-dioxa-3H-perfluorononanoic acid (ADONA)	ND	2.0		ng/L	1	EPA 537.1	11/22/19	11/27/19 4:24	BLM	
Surrogates		% Recovery	Recovery Limits		Flag/Qual					
13C-PFHxA		102	70-130					11/27/19 4:24		
M3HFPO-DA		96.5	70-130					11/27/19 4:24		
13C-PFDA		123	70-130					11/27/19 4:24		
d5-NEtFOSAA		125	70-130					11/27/19 4:24		

39 Spruce Street * East Longmeadow, MA 01028 * FAX 413/525-6405 * TEL. 413/525-2332

Sample Extraction Data

Prep Method: EPA 537-EPA 537.1

Lab Number [Field ID]	Batch	Initial [mL]	Final [mL]	Date
19K0630-01 [Well #1 Field Blank]	B246729	250	1.00	11/22/19
19K0630-02 [Well #1]	B246729	250	1.00	11/22/19

39 Spruce Street * East Longmeadow, MA 01028 * FAX 413/525-6405 * TEL. 413/525-2332

QUALITY CONTROL

Semivolatile Organic Compounds by - LC/MS-MS - Quality Control

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
Batch B246729 - EPA 537										
Blank (B246729-BLK1)										
Prepared: 11/22/19 Analyzed: 11/27/19										
Perfluorobutanesulfonic acid (PFBS)	ND	2.0	ng/L							
Perfluorohexanoic acid (PFHxA)	ND	2.0	ng/L							
Perfluorohexanesulfonic acid (PFHxS)	ND	2.0	ng/L							
Perfluoroheptanoic acid (PFHpA)	ND	2.0	ng/L							
Perfluorooctanoic acid (PFOA)	ND	2.0	ng/L							
Perfluorooctanesulfonic acid (PFOS)	ND	2.0	ng/L							
Perfluorononanoic acid (PFNA)	ND	2.0	ng/L							
Perfluorodecanoic acid (PFDA)	ND	2.0	ng/L							
N-EtFOSAA	ND	2.0	ng/L							
Perfluoroundecanoic acid (PFUnA)	ND	2.0	ng/L							
N-MeFOSAA	ND	2.0	ng/L							
Perfluorododecanoic acid (PFDoA)	ND	2.0	ng/L							
Perfluorotridecanoic acid (PFTTrDA)	ND	2.0	ng/L							
Perfluorotetradecanoic acid (PFTA)	ND	2.0	ng/L							
Hexafluoropropylene oxide dimer acid (HFPO-DA)	ND	2.0	ng/L							
11Cl-PF3OUdS (F53B Major)	ND	2.0	ng/L							
9Cl-PF3ONS (F53B Minor)	ND	2.0	ng/L							
4,8-dioxa-3H-perfluorononanoic acid (ADONA)	ND	2.0	ng/L							
Surrogate: 13C-PFHxA	45.8		ng/L	40.0		115	70-130			
Surrogate: M3HFPO-DA	45.4		ng/L	40.0		114	70-130			
Surrogate: 13C-PFDA	43.7		ng/L	40.0		109	70-130			
Surrogate: d5-NEtFOSAA	167		ng/L	160		104	70-130			
LCS (B246729-BS1)										
Prepared: 11/22/19 Analyzed: 11/27/19										
Perfluorobutanesulfonic acid (PFBS)	9.44	2.0	ng/L	10.0		94.4	70-130			
Perfluorohexanoic acid (PFHxA)	10.3	2.0	ng/L	10.0		103	70-130			
Perfluorohexanesulfonic acid (PFHxS)	9.89	2.0	ng/L	9.10		109	70-130			
Perfluoroheptanoic acid (PFHpA)	10.3	2.0	ng/L	10.0		103	70-130			
Perfluorooctanoic acid (PFOA)	10.9	2.0	ng/L	10.0		109	70-130			
Perfluorooctanesulfonic acid (PFOS)	10.2	2.0	ng/L	9.25		110	70-130			
Perfluorononanoic acid (PFNA)	11.5	2.0	ng/L	10.0		115	70-130			
Perfluorodecanoic acid (PFDA)	12.0	2.0	ng/L	10.0		120	70-130			
N-EtFOSAA	12.1	2.0	ng/L	10.0		121	70-130			
Perfluoroundecanoic acid (PFUnA)	11.6	2.0	ng/L	10.0		116	70-130			
N-MeFOSAA	9.95	2.0	ng/L	10.0		99.5	70-130			
Perfluorododecanoic acid (PFDoA)	10.8	2.0	ng/L	10.0		108	70-130			
Perfluorotridecanoic acid (PFTTrDA)	10.8	2.0	ng/L	10.0		108	70-130			
Perfluorotetradecanoic acid (PFTA)	10.4	2.0	ng/L	10.0		104	70-130			
11Cl-PF3OUdS (F53B Major)	8.73	2.0	ng/L	9.40		92.9	70-130			
9Cl-PF3ONS (F53B Minor)	8.88	2.0	ng/L	9.30		95.5	70-130			
4,8-dioxa-3H-perfluorononanoic acid (ADONA)	9.94	2.0	ng/L	10.0		99.4	70-130			
Surrogate: 13C-PFHxA	40.7		ng/L	40.0		102	70-130			
Surrogate: M3HFPO-DA	39.6		ng/L	40.0		99.0	70-130			
Surrogate: 13C-PFDA	45.2		ng/L	40.0		113	70-130			
Surrogate: d5-NEtFOSAA	174		ng/L	160		109	70-130			

39 Spruce Street * East Longmeadow, MA 01028 * FAX 413/525-6405 * TEL. 413/525-2332

FLAG/QUALIFIER SUMMARY

*	QC result is outside of established limits.
†	Wide recovery limits established for difficult compound.
‡	Wide RPD limits established for difficult compound.
#	Data exceeded client recommended or regulatory level
ND	Not Detected
RL	Reporting Limit is at the level of quantitation (LOQ)
DL	Detection Limit is the lower limit of detection determined by the MDL study
MCL	Maximum Contaminant Level

Percent recoveries and relative percent differences (RPDs) are determined by the software using values in the calculation which have not been rounded.

No results have been blank subtracted unless specified in the case narrative section.

CERTIFICATIONS

Certified Analyses included in this Report

Analyte	Certifications
<i>EPA 537.1 in Drinking Water</i>	
Perfluorobutanesulfonic acid (PFBS)	NH-P,VT-DW,NJ,CT
Perfluorohexanoic acid (PFHxA)	NH-P,VT-DW,NJ,CT
Perfluorohexanesulfonic acid (PFHxS)	NH-P,VT-DW,NJ,CT
Perfluoroheptanoic acid (PFHpA)	NH-P,VT-DW,NJ,CT
Perfluorooctanoic acid (PFOA)	VT-DW,NJ,NY,NH,CT
Perfluorooctanesulfonic acid (PFOS)	VT-DW,NJ,NY,NH,CT
Perfluorononanoic acid (PFNA)	NH-P,VT-DW,NJ,CT
Perfluorodecanoic acid (PFDA)	NH-P,VT-DW,NJ,CT
N-EtFOSAA	NH-P,VT-DW,NJ,CT
Perfluoroundecanoic acid (PFUnA)	NH-P,VT-DW,NJ,CT
N-MeFOSAA	NH-P,VT-DW,NJ,CT
Perfluorododecanoic acid (PFDoA)	NH-P,VT-DW,NJ,CT
Perfluorotridecanoic acid (PFTrDA)	NH-P,VT-DW,NJ,CT
Perfluorotetradecanoic acid (PFTA)	NH-P,VT-DW,NJ,CT
Hexafluoropropylene oxide dimer acid (HFPO-DA)	NH-P,VT-DW,NJ,CT
11Cl-PF3OUdS (F53B Major)	NH-P,VT-DW,NJ,CT
9Cl-PF3ONS (F53B Minor)	NH-P,VT-DW,NJ,CT
4,8-dioxa-3H-perfluorononanoic acid (ADONA)	NH-P,VT-DW,NJ,CT

The CON-TEST Environmental Laboratory operates under the following certifications and accreditations:

Code	Description	Number	Expires
AIHA	AIHA-LAP, LLC - ISO17025:2017	100033	03/1/2020
MA	Massachusetts DEP	M-MA100	06/30/2020
CT	Connecticut Department of Public Health	PH-0567	09/30/2021
NY	New York State Department of Health	10899 NELAP	04/1/2020
NH-S	New Hampshire Environmental Lab	2516 NELAP	02/5/2020
RI	Rhode Island Department of Health	LAO00112	12/30/2019
NC	North Carolina Div. of Water Quality	652	12/31/2019
NJ	New Jersey DEP	MA007 NELAP	06/30/2020
FL	Florida Department of Health	E871027 NELAP	06/30/2020
VT	Vermont Department of Health Lead Laboratory	LL015036	07/30/2020
ME	State of Maine	2011028	06/9/2021
VA	Commonwealth of Virginia	460217	12/14/2019
NH-P	New Hampshire Environmental Lab	2557 NELAP	09/6/2020
VT-DW	Vermont Department of Health Drinking Water	VT-255716	06/12/2020
NC-DW	North Carolina Department of Health	25703	07/31/2020
PA	Commonwealth of Pennsylvania DEP	68-05812	06/30/2020



Phone: 413-525-2332

Fax: 413-525-6405

Email: info@contestlabs.com

http://www.contestlabs.com

CHAIN OF CUSTODY RECORD

39 Spruce Street
East Longmeadow, MA 01028

Doc # 381 Rev 2_06262019

Page ____ of ____

Eastern Water Solutions
 Address: 5 Beaver Rd Oxford CT 06478
 Phone: 203-264-5155
 Project Location: 4 Barbourtown Rd Canton CT 06019
 Project Number: Well # 1 1/2 hp CT 03039 2
 Project Manager: Alicia Gaffin / Shane Grant
 Con-Test Quote Name/Number: Well # 1 1/2 hp
 Invoice Recipient:
 Sampled By: Alicia Gaffin

Requested Turnaround Time
 7-Day 10-Day Field Filtered
 PFAS 10-Day (std) Due Date: ASAP Lab to Filter

Rush-Approval Required
 1-Day 3-Day Field Filtered
 2-Day 4-Day Lab to Filter

Data Delivery
 Format: PDF EXCEL
 Other:
 CLP Like Data Pkg Required:
 Email To: Alicia@easternwater.com
 Fax To #: Shane@easternwater.com

ANALYSIS REQUESTED																					
VIALS	GLASS	PLASTIC	BACTERIA	ENCORE																	

Field Blak
PFAS
PFAS

² Preservation Code
 Courier Use Only
 Total Number Of:
 VIALS _____
 GLASS _____
 PLASTIC _____
 BACTERIA _____
 ENCORE _____

Glassware in the fridge? Y / N
 Glassware in freezer? Y / N
 Prepackaged Cooler? Y / N

*Contest is not responsible for missing samples from prepacked coolers

Con-Test Work Order#	Client Sample ID / Description	Beginning Date/Time	Ending Date/Time	COMP/GRAB	Matrix Code	Conc Code	VIALS	GLASS	PLASTIC	BACTERIA	ENCORE
1	Well # 1 Field Blak	11/11/19	12 pm		G				1		
2	Well # 1 1/2 hp	11/11/19	12:00 pm		DW				1		
3	Well # 1 1/2 hp	11/11/19	12:10 pm		DW				1		

¹ Matrix Codes:
 GW = Ground Water
 WW = Waste Water
 DW = Drinking Water
 A = Air
 S = Soil
 SL = Sludge
 SOL = Solid
 O = Other (please define)

Relinquished by: (signature) _____ Date/Time: 11/11/19 2:13
 Received by: (signature) _____ Date/Time: 11/11/19 1:22
 Relinquished by: (signature) _____ Date/Time: _____
 Received by: (signature) _____ Date/Time: _____
 Relinquished by: (signature) _____ Date/Time: _____
 Received by: (signature) _____ Date/Time: _____
 Relinquished by: (signature) _____ Date/Time: _____
 Received by: (signature) _____ Date/Time: _____

Client Comments: Including PFAS in EPA Method 537 Revision

Per client - submitted 3 bottles as rcv'd 3 bottles with the kit, so wrote all 3 on COC but only need to run the FB and the 1 bottle - the 3rd is extra if needed 11/12/19 mmk

MCP Certification Form Required
 CT RCP Required
 RCP Certification Form Required
 MA State DW Required
 PWSID # _____

Indicate possible sample concentration within the Conc Code column above:
 H - High; M - Medium; L - Low; C - Clean; U - Unknown

Other _____

Project Entity
 Government Municipality MWRA WRTA
 Federal 21 J School
 City Brownfield MBTA

Other Chromatogram
 PCB ONLY Soxhlet
 AIHA-LAP, LLC Non Soxhlet

Comments:

Disclaimer: Con-Test Labs is not responsible for any omitted information on the Chain of Custody. The Chain of Custody is a legal document that must be complete and accurate and is used to determine what analyses the laboratory will perform. Any missing information is not the laboratory's responsibility. Con-Test values your partnership on each project and will try to assist with missing information, but will not be held accountable.

I Have Not Confirmed Sample Container Numbers With Lab Staff Before Relinquishing Over Samples _____



con-test
ANALYTICAL LABORATORY

Doc# 277 Rev 5 2017

Login Sample Receipt Checklist - (Rejection Criteria Listing - Using Acceptance Policy) Any False Statement will be brought to the attention of the Client - State True or False

Client Eastern Water Solutions

Received By SA Date 11/11 Time 1422

How were the samples received? In Cooler T No Cooler _____ On Ice T No Ice _____
Direct from Sampling _____ Ambient _____ Melted Ice _____

Were samples within Temperature? 2-6°C _____ By Gun # 5 Actual Temp - 2.8
By Blank # _____ Actual Temp - _____

Was Custody Seal Intact? NA Were Samples Tampered with? NA

Was COC Relinquished? T Does Chain Agree With Samples? T

Are there broken/leaking/loose caps on any samples? F

Is COC in ink/ Legible? _____ Were samples received within holding time? T

Did COC include all pertinent Information? Client T Analysis T Sampler Name T
Project T ID's T Collection Dates/Times T

Are Sample labels filled out and legible? T

Are there Lab to Filters? F Who was notified? _____

Are there Rushes? F Who was notified? _____

Are there Short Holds? F Who was notified? _____

Is there enough Volume? T

Is there Headspace where applicable? NA

Proper Media/Containers Used? T MS/MSD? F

Were trip blanks received? F Is splitting samples required? F

Do all samples have the proper pH? _____ On COC? f Acid NA Base NA

Vials	#	Containers:	#	#	#
Unp-		1 Liter Amb.		1 Liter Plastic	16 oz Amb.
HCL-		500 mL Amb.		500 mL Plastic	8oz Amb/Clear
Meoh-		250 mL Amb.		250 mL Plastic	4oz Amb/Clear
Bisulfate-		Flashpoint		Col./Bacteria	2oz Amb/Clear
DI-		Other Glass		Other Plastic	Encore
Thiosulfate-		SOC Kit		Plastic Bag	Frozen:
Sulfuric-		Perchlorate		Ziplock	

Unused Media

Vials	#	Containers:	#	#	#
Unp-		1 Liter Amb.		1 Liter Plastic	16 oz Amb.
HCL-		500 mL Amb.		500 mL Plastic	8oz Amb/Clear
Meoh-		250 mL Amb.		250 mL Plastic	4oz Amb/Clear
Bisulfate-		Col./Bacteria		Flashpoint	2oz Amb/Clear
DI-		Other Plastic		Other Glass	Encore
Thiosulfate-		SOC Kit		Plastic Bag	Frozen:
Sulfuric-		Perchlorate		Ziplock	

Comments: